



NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-A,B

Lot No. 62014

Source of Material: Monoclonal antibody

Mfg. Date 18/12/2019

Volume 10 mL

Exp. Date 18/06/2021

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed red No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 – 7.50	Complied
4. Titer	A ₁ cell ≥ 1:256 A ₂ cell ≥ 1:128 B cell ≥ 1:256	Complied
5. Specificity test	A ₁ cell = 4+ A ₂ cell = 4+ B cell = 4+ O cell = Negative	Complied
6. Avidity test	A ₁ cell < 6 sec A ₂ cell < 20 sec B cell < 6 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Thiti B.

Date 10/07/2563

(Thitiphorn Bhakbhumpung)

Quality Control Manager

National Blood Centre, Thai Red Cross Society