



NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-A,B

Lot No. 63011

Source of Material: Monoclonal antibody

Mfg. Date 19 AUG 20

Volume 10 mL, 50 mL

Exp. Date 19 FEB 22

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed red No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 – 7.50	Complied
4. Titer	A ₁ cell ≥ 1:256 A ₂ cell ≥ 1:128 B cell ≥ 1:256	Complied
5. Specificity test	A ₁ cell = 4+ A ₂ cell = 4+ B cell = 4+ O cell = Negative	Complied
6. Avidity test	A ₁ cell < 6 sec A ₂ cell < 20 sec B cell < 6 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Thiti B. Date 01/10/2563

(Thitiphorn Bhakbhumpong)

Quality Control Manager

National Blood Centre, Thai Red Cross Society