



ศูนย์บริการโลหิตแห่งชาติ
สภากาชาดไทย

NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-B

Lot No. 63015

Source of Material: Monoclonal antibody

Mfg. Date 30 MAR 20

Volume 10 mL

Exp. Date 30 SEP 21

| Test Items | Specifications | Results |
|--------------------------|--|----------|
| 1. Appearance | Clear liquid dyed yellow No turbidity, precipitate, particle or gel formation | Complied |
| 2. Label | Complete, Correct and Clear | Complied |
| 3. pH | 6.50 – 7.50 | Complied |
| 4. Titer | B cell \geq 1:256 A ₁ B cell \geq 1:128 | Complied |
| 5. Specificity test | B cell = 4+ A ₁ B cell = 4+ A ₁ cell = Negative O cell = Negative | Complied |
| 6. Avidity test | B cell < 6 sec A ₁ B cell < 6 sec | Complied |
| 7. Microbial enumeration | NMT 1.0 CFU/mL | Complied |

Thiti-B.

Date 28/08/2563

(Thitiphorn Bhakbhumpong)

Quality Control Manager

National Blood Centre, Thai Red Cross Society