



# NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

## CERTIFICATE OF ANALYSIS

Product: Anti-B

Lot No. 63031

Source of Material: Monoclonal antibody

Mfg. Date 30 DEC 20

Volume 10 mL

Exp. Date 30 JUN 22

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed yellow No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 – 7.50	Complied
4. Titer	B cell $\geq$ 1:256 A <sub>1</sub> B cell $\geq$ 1:128	Complied
5. Specificity test	B cell = 4+ A <sub>1</sub> B cell = 4+ A <sub>1</sub> cell = Negative O cell = Negative	Complied
6. Avidity test	B cell < 6 sec A <sub>1</sub> B cell < 6 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Thiti B.

Date 03/02/2564

(Thitiphorn Bhakbhumpong)

Quality Control Manager

National Blood Centre, Thai Red Cross Society