



NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-B Lot No. 63032
Source of Material: Monoclonal antibody Mfg. Date 30 DEC 20
Volume 10 mL Exp. Date 30 JUN 22

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed yellow No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 – 7.50	Complied
4. Titer	B cell \geq 1:256 A ₁ B cell \geq 1:128	Complied
5. Specificity test	B cell = 4+ A ₁ B cell = 4+ A ₁ cell = Negative O cell = Negative	Complied
6. Avidity test	B cell < 6 sec A ₁ B cell < 6 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Thiti B Date 04/03/2564

(Thitiphorn Bhakbhumpong)

Quality Control Manager

National Blood Centre, Thai Red Cross Society