



NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-A

Lot No. 63025

Source of Material: Monoclonal antibody

Mfg. Date 13 MAY 20

Volume 10 mL

Exp. Date 13 NOV 21

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed blue No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 – 7.50	Complied
4. Titer	A ₁ cell ≥ 1:256 A ₂ B cell ≥ 1:64	Complied
5. Specificity test	A ₁ cell = 4+ A ₂ B cell >2+ B cell = Negative O cell = Negative	Complied
6. Avidity test	A ₁ cell < 6 sec A ₂ B cell < 30 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Thiti B.

Date 01/10/2563

(Thitiphorn Bhakbhumpong)

Quality Control Manager

National Blood Centre, Thai Red Cross Society